

FACSIMILE COVER SHEET

Date:

4-27-01

To:

Fax No:

From:

Tel. No:

EX M. CHOBBIN GAU 2621

703-306-5406

W.A. WEBB

312-321-4200

Client No:

5050

No. of Pages

(inc. this page):

5

Confirmation Copy To Follow: Yes ☐No ☒IF YOU HAVE ANY PROBLEMS RECEIVING THIS MESSAGE,
PLEASE CALL 312-321-4200 AND ASK FOR: _____

THIS MESSAGE IS INTENDED ONLY FOR THE INDIVIDUAL OR ENTITY TO WHICH IT IS ADDRESSED. IT MAY CONTAIN PRIVILEGED, CONFIDENTIAL, ATTORNEY WORK PRODUCT, OR TRADE SECRET INFORMATION WHICH IS EXEMPT FROM DISCLOSURE UNDER APPLICABLE LAWS. IF YOU ARE NOT THE INTENDED RECIPIENT, OR AN EMPLOYEE OR AGENT RESPONSIBLE FOR DELIVERING THE MESSAGE TO THE INTENDED RECIPIENT, YOU ARE HEREBY NOTIFIED THAT ANY DISSEMINATION, DISTRIBUTION, OR COPYING OF THIS MESSAGE IS STRICTLY PROHIBITED. IF YOU HAVE RECEIVED THIS MESSAGE IN ERROR, PLEASE NOTIFY US IMMEDIATELY BY TELEPHONE AND RETURN THE ORIGINAL MESSAGE (AND ALL COPIES) TO US BY MAIL AT THE ABOVE ADDRESS. WE WILL REIMBURSE YOU FOR POSTAGE.

COVER MESSAGE:

Re USSN 09/369,510

Mr CHOBBIN -

PLEASE CALL TO SCHEDULE THE
TELEPHONE INTERVIEW.

W.A. WEBB

312-321-4218

**BRINKS
HOFER
GILSON
& LIONE**A PROFESSIONAL CORPORATION
INTELLECTUAL PROPERTY ATTORNEYSNBC Tower - Suite 3800
435 N. CITYFRONT PLAZA DRIVE
CHICAGO, ILLINOIS 60611-6599
FACSIMILE 312-321-4200
TELEPHONE 312-321-4200
E-MAIL wbgil@mcsl.com

Our Case No. 5050/584-1**IN THE UNITED STATES PATENT AND TRADEMARK OFFICE**

In re Application of:)	
John A. Hossack et al.)	
Serial No. 09/369,510)	Examiner: M. Choobin
Filing Date: August 6, 1999)	Group Art Unit No. 2621
For Medical Diagnostic Ultrasonic)	
Imaging Method and System for)	
Displaying Multi-Frame Images)	

AGENDA FOR TELEPHONE INTERVIEW**A. Review the preferred embodiment**

1. Multiple frames are acquired, and each frame is identified with a particular phase of a cycle (e.g., the cardiac cycle). For example, each frame may be labeled with one of ten separate phases, and ten frames may be collected for each transducer position (See p. 2, l. 26 - p. 3, l. 6).
2. Multiple images are generated from the frames of (1), and each image uses data from multiple frames, all of about the same phase. Examples are images constructed from a 3-D data set (p. 3, l. 25 - p. 4, l. 3 and p. 6, ll. 2-4) and extended field of view images (p. 6, ll. 10-16 and 28-30).

B. Review Gandini U.S. Pat. 5,645,066

1. The frames used to create 3-D images are not selected based on cycle (e.g. cardiac) phase (See col. 4, ll. 6-10 and col. 6, ll. 22-49, which call for no selection of frames by cycle phase).

Agenda for Telephone Interview
Page 2 of 4

2. The Cine-loop description at col. 4, ll. 54 relates to 2-D image sequences which show tissue motion, not to 3-D images or to extended field of view images.

C. Compare preferred embodiment with Gandini

1. Figure A shows an example of frames used to form 3-D data set or extended field of view image in a preferred embodiment.
2. Figure B shows an example of frames used to form 3-D data set in Gandini.
3. Gandini does not recognize the problem solved by this invention.

D. Review Claims 24 and 32

1. Claim 24(b) and (c)
2. Claim 32(b) and (c)
3. Proposed amendments
 32. (Amended) An ultrasonic imaging system comprising:
 - (a) means for acquiring image data for a plurality of frames, each frame identified with a respective phase of a physiological cycle;
 - (b) means for generating a first image associated with a first phase of the physiological cycle from image data from [at least two] multiple selected ones of the frames of (a) associated with [a] the first phase of the physiological cycle;
 - (c) means for generating a second image associated with a second phase of the physiological cycle from image data from [at least two] multiple selected ones of the frames of (a) associated with [a] the second phase of the physiological cycle; and

✓ Agenda for Telephone Interview
Page 3 of 4


(d) means for displaying at least the first and second images in sequence to a user.

E. Review the rejection of claims 29 and 36


1. The specification provides support for these claims at page 6, line 13 and in Figure 6.

Fig A

A1-A5 ARE ALL 0° PHASE, AND IMAGE A IS
CONSTRUCTED FROM SELECTED ONES OF A1-A5.



B1-B5 ARE ALL 180° PHASE, AND IMAGE B IS
CONSTRUCTED FROM SELECTED ONES OF B1-B5.

Fig B (GANDINI)

GANDINI 3-D IMAGES ARE CREATED FROM SOME
OR ALL OF FRAMES 1-10, WITHOUT SORTING THE
FRAMES BY PHASE ANGLE.